

A miniature oscillating microbalance for sampling ice and volcanic ash from a small airborne platform

Article

Published Version

Creative Commons: Attribution 4.0 (CC-BY)

Open access

Airey, M. ORCID: <https://orcid.org/0000-0002-9784-0043>,
Harrison, G. ORCID: <https://orcid.org/0000-0003-0693-347X>,
Nicoll, K. ORCID: <https://orcid.org/0000-0001-5580-6325>,
Williams, P. ORCID: <https://orcid.org/0000-0002-9713-9820>
and Marlton, G. (2017) A miniature oscillating microbalance for
sampling ice and volcanic ash from a small airborne platform.
Review of Scientific Instruments, 88 (8). 086108. ISSN 0034-
6748 doi: <https://doi.org/10.1063/1.4998971> Available at
<https://centaur.reading.ac.uk/71788/>

It is advisable to refer to the publisher's version if you intend to cite from the work. See [Guidance on citing](#).

To link to this article DOI: <http://dx.doi.org/10.1063/1.4998971>

Publisher: AIP

All outputs in CentAUR are protected by Intellectual Property Rights law, including copyright law. Copyright and IPR is retained by the creators or other copyright holders. Terms and conditions for use of this material are defined in the [End User Agreement](#).

www.reading.ac.uk/centaur

CentAUR

Central Archive at the University of Reading

Reading's research outputs online

Note: A miniature oscillating microbalance for sampling ice and volcanic ash from a small airborne platform

M. W. Airey, R. G. Harrison, K. A. Nicoll, P. D. Williams, and G. J. Marlton

Citation: [Review of Scientific Instruments](#) **88**, 086108 (2017); doi: 10.1063/1.4998971

View online: <http://dx.doi.org/10.1063/1.4998971>

View Table of Contents: <http://aip.scitation.org/toc/rsi/88/8>

Published by the [American Institute of Physics](#)

Articles you may be interested in

[Temperature-stabilized differential amplifier for low-noise DC measurements](#)

[Review of Scientific Instruments](#) **88**, 085106 (2017); 10.1063/1.4997963

[TinyLev: A multi-emitter single-axis acoustic levitator](#)

[Review of Scientific Instruments](#) **88**, 085105 (2017); 10.1063/1.4989995

[A novel interferometric method for the study of the viscoelastic properties of ultra-thin polymer films determined from nanobubble inflation](#)

[Review of Scientific Instruments](#) **88**, 093901 (2017); 10.1063/1.5000948

[Real-time digital signal recovery for a multi-pole low-pass transfer function system](#)

[Review of Scientific Instruments](#) **88**, 085104 (2017); 10.1063/1.4990810

[The world's smallest capacitive dilatometer, for high-resolution thermal expansion and magnetostriction in high magnetic fields](#)

[Review of Scientific Instruments](#) **88**, 083903 (2017); 10.1063/1.4997073

[Note: A high-frequency signal generator based on direct digital synthesizer and field-programmable gate array](#)

[Review of Scientific Instruments](#) **88**, 096103 (2017); 10.1063/1.5001489



Obstruction free access
optical table with integrated cryocooler



Various Objective Options

attoDRY800

- Cryogenic Temperatures
- Ultra-Low Vibration
- Optical Table Included
- Fast Cooldown



5% DISCOUNT

on all nanopositioners purchased
for your attoDRY800 set up*
Coupon Code: PTJAD800

*valid for quotations issued before November, 2017

Note: A miniature oscillating microbalance for sampling ice and volcanic ash from a small airborne platform

M. W. Airey, R. G. Harrison, K. A. Nicoll, P. D. Williams, and G. J. Marlon
Department of Meteorology, University of Reading, Reading RG6 6BB, United Kingdom

(Received 19 May 2017; accepted 30 July 2017; published online 11 August 2017)

A lightweight and low power oscillating microbalance for *in situ* sampling of atmospheric ice and volcanic ash is described for airborne platforms. Using a freely exposed collecting wire fixed at only one end to a piezo transducer, the instrument collects airborne materials. Accumulated mass is determined from the change in natural frequency of the wire. The piezo transducer is used in a dual mode to both drive and detect the oscillation. Three independent frequency measurement techniques are implemented with an on-board microcontroller: a frequency sweep, a Fourier spectral method, and a phase-locked loop. These showed agreement to ± 0.3 Hz for a 0.5 mm diameter collecting wire of 120 mm long, flown to 19 km altitude on a weather balloon. The instrument is well suited to disposable use with meteorological radiosondes, to provide high resolution vertical profiles of mass concentration. © 2017 Author(s). All article content, except where otherwise noted, is licensed under a Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>). [<http://dx.doi.org/10.1063/1.4998971>]

In extreme situations posing hazards to aircraft, such as after volcanic eruptions, safety-critical decisions require the distribution and concentration of hazardous materials to be known. Remote sensing techniques are of limited value for this as they cannot retrieve mass concentrations without prior knowledge of ash properties, such as refractive index.^{1,2} An alternative approach is *in situ* measurement, and the instrument described here measures the mass directly. Beyond volcanic ash, other applications include detection of ice accretion within supercooled liquid water (SLW) clouds. Our instrument is compact, disposable, lightweight (40 g), and low-power (~20 mA from a 9 V supply) designed for rapid-response deployment using a standard weather balloon or other (un)manned airborne platforms.

This instrument provides a direct mass determination using the established oscillating microbalance approach, with a wire collector.³ Unlike previous implementations which restricted the airflow by actuating devices and enclosures,⁴ it uses a freely exposed collecting wire. This both improves the sampling efficiency and yields a response more accurately following theory of a vibrating cylinder fixed at one end. A piezo-electric transducer at the fixed end is used to both drive and detect the oscillation, removing the need for mechanical actuation and associated power consumption and weight. Frequency measurement is implemented through three independent techniques to increase confidence in the measurement. The device is coupled to a meteorological radiosonde which also provides power and data telemetry, through a PANDORA (Programmable ANalogue and Digital Operational Radiosonde Accessory) interface.⁵ For the ash collection application, the collecting wire is coated with high-tack adhesive (gasket glue), which can be evenly applied, is effective at cold temperatures, and has been found in laboratory tests to effectively retain ash particles over a range of grain sizes from 10s μm to mm scale.

The collecting wire on our instrument, of 120 mm length and 0.5 mm diameter, is soldered to a piezo transducer with

feedback monitoring capability, in contrast to the previous design,³ which utilised a drive coil at the attached end of the wire and a pickup coil at the free end. For this new design, the transducer is used in a dual mode, both to excite oscillations and to measure the oscillations established via a feedback signal. The combination of the wire sensor, a phase-locked loop, and programmable microcontroller, which is illustrated in Fig. 1, allows for a great versatility of use that may be tailored to specific uses. Three combined methods of wire vibration frequency detection may be programmed and are described below: direct period measurement via a phase-locked loop (PLL method), a spectral Fast Hartley Transform technique (FHT method), and observation of peak amplitude at driven frequencies (SWEEP method).

In use on an airborne platform, the data gathered and processed on the instrument microcontroller are transferred to the PANDORA control board over the I²C protocol at 1 Hz. Four two-byte “word” variables, the oscillation amplitude and the three independently determined frequency measurements, are transferred during each data request. These data are transmitted with the standard radiosonde data stream and can be retrieved with a commercial radiosonde base station.

The microcontroller used in this device (ATMega 328) has a 32 kB flash memory and a 10 bit analogue to digital convertor (ADC), allowing relatively complex code to be implemented. For the driven operations, timed digital pulses on one of the microcontroller’s output pins send a square wave at a defined frequency to the piezo transducer, exciting the wire at that frequency. The vibration of the wire, whether driven or natural, generates a fluctuating voltage at the piezo transducer. This is amplified for the 5 V range of the ADC and is digitised by the microcontroller. The oscillation voltage is also processed through a further gain stage before being applied to a phase-locked loop (PLL), implemented with an HEF4046 integrated circuit. When the PLL is locked to the input signal, a lock indication is provided to trigger the frequency measurement




FIG. 1. (a) Sketch of the oscillating microbalance collector. (b) Block diagram to illustrate the main components and signal flow. Dashed lines show optional drive provided to the piezo transducer, dotted lines show the PLL “lock” state signal, and solid lines show the signal received from the wire vibration. White boxes indicate hardware and grey boxes indicate software. Signal flow and component definitions are described in the text.

by the microcontroller. The PLL time constant is set to stabilise the input frequencies.

The SWEEP method is achieved by first driving the wire for short bursts (~ 1 s) over a range of frequencies at 1 Hz steps and measuring the amplitude $300\ \mu\text{s}$ after each burst has ceased. Successively narrower sweeps (0.5 and 0.1 Hz steps) are then centred upon that peak value obtained to refine the measurement. The finest sweep may then be repeated for as long as measurements are required. This method offers poor temporal resolution (it is slow, with even the finest sweeps taking ~ 10 to 15 s over reasonable ranges of frequencies) but is capable of determining the natural frequency to a precision of ± 0.05 Hz. Figure 2 shows an example amplitude response to a long duration fine sweep (0.1 Hz steps) for a steel wire of 120 mm length, producing a maximum amplitude at 26 Hz. The wire was subsequently coated uniformly with an additional 200 mg of cosmetic nitrocellulose to increase its mass (roughly the same mass as the unladen wire); the dashed line shows the changed amplitude response of the wire, with a new maximum amplitude at 20.5 Hz.

The PLL method measures the frequency directly from the signal. The PLL output indicates the “lock” state (high or low) from filtering the phase-lock pulses, which identifies when the wire is vibrating at its natural frequency as described above. When the “lock” state is high, the microcontroller performs a frequency measurement using a reciprocal timing method. Each measurement is derived from the mean of five periods to reduce the sampling error. This method is fast, providing continually updated readings at a rate corresponding to the data transfer rate (1 Hz) while the lock state is high. Long duration runs of this method have found a 3σ error of $\sim \pm 0.2$ Hz.

The FHT method uses the Hartley transform,⁶ which is a form of Fourier signal processing that uses only real input and output values to avoid the computation of complex

numbers. This makes it less computationally expensive for implementation on a small microprocessor. Here, the fast Hartley transform⁷ is used to determine the natural frequency of the wire by using the amplitude input taken from the first gain stage; the Hann window is applied to the spectrum during the calculations. This algorithm can be modified to take up to 256 samples from the amplitude input at any specified sampling frequency. An optimum arrangement for this system is found to be 256 samples at 100 Hz, which provides a spectrum over the frequency range 0 to 50 Hz with ~ 0.4 Hz bin size. This procedure is much faster than the SWEEP method described




FIG. 2. Relative amplitude frequency response of a 120 mm long, 0.5 mm diameter steel wire after being driven for short pulse intervals (~ 1 s) at 0.1 Hz steps over a range of frequencies. The solid line is for an unladen wire and the dashed line is for the same wire loaded evenly with an additional 200 mg of material.

above, taking only ~ 2.5 s to compute the full 50 Hz range. However, there is a trade-off in terms of resolution, as the bin size is ± 0.4 Hz. Figure 3 shows an example spectrum from a vibrating wire with an independently determined natural frequency (using the SWEEP method) of 26 Hz. The ten bins at either extreme of the data range are not counted, resulting in an effective range of ~ 4 to ~ 46 Hz. The frequency range used and initial frequency defined (f_0) were chosen to maximize the output resolution whilst mitigating the boundary effects of the transform, for a practical wire length.

By combining these techniques, it is possible to specify an optimum system in both temporal resolution and measurement precision. The three methods are always found to agree within their limits of uncertainty. The procedure to maximize speed and efficiency is to initially execute an FHT frequency measurement to define a target region on which to apply sweeps for increased precision without the need for long initial coarser runs. The PLL frequency method is employed simultaneously to provide increased data acquisition resolution. Parallel data acquisition has the added advantage of mitigating the risk of data loss. The flexibility inherent in these alternative approaches means that the instrument can be tailored to a specific task, e.g., a high temporal resolution with low precision strategy may be more suited to a task than a low temporal resolution with high precision strategy.

To demonstrate the mass retrieval, the frequency at the peak amplitude was measured as the accreted mass was increased. Uniform (20 ± 5 mg) coatings of cosmetic nitrocellulose were consecutively added to the wire. This material allowed consistent mass increments to a higher precision than that afforded by using a natural material such as ash. Figure 4 shows that the response can be represented by a least square linear fit, giving a sensitivity of -0.024 Hz mg^{-1} (-40.6 mg Hz^{-1}), with an R^2 of 0.99. The standard error of the frequency and mass are, respectively, 0.27 Hz and 10.9 mg. Using this response, the associated mass concentration may be calculated through the use of the collection efficiency⁸ of the wire.




FIG. 3. Histogram of relative bin magnitudes resulting from a 256 point FHT with Hann windowing, sampled at a rate of 100 Hz and grouped into bins of ~ 0.4 Hz for a wire vibrating at a natural frequency of 26 Hz. Dotted bars are for discarded bins and are not counted in the analysis.




FIG. 4. Response of the 0.5 mm vibrating wire's natural frequency to an incremental mass increase. A linear fit indicates a sensitivity of -0.02 Hz mg^{-1} with a standard error of 0.27 Hz and an R^2 of 0.99. Vertical error bars represent the frequency measurement precision of ± 0.05 Hz and horizontal error bars represent the mass measurement precision of ± 5 mg.

In summary, a novel, highly versatile oscillating microbalance has been developed to collect atmospheric material. The sensor wire has a measurement range of >300 mg and the *in situ* operational effectiveness has been proven in test balloon flights. Data acquired from the instrument during a balloon deployment at Reading to 19 km altitude in clear conditions provided frequency measurements from all three methods, whose mean profiles agreed to ± 0.3 Hz. It has also successfully collected and quantified suspended volcanic ash in the laboratory. The combined use of three independent frequency measurements facilitates tailored use to suit a variety of applications, including measurements of volcanic ash concentration, SLW concentration via ice accretion, and Saharan dust.

This work was funded by the UK Natural Environment Research Council (NERC), Grant No. NE/P003362/1 (VOLCLAB). K.A.N. acknowledges the support of NERC through Independent Research Fellowship No. NE/L011514/1. P.D.W. is funded by The Royal Society, Grant No. UF130571. The authors thank Stefan Kneifel of the Institute of Geophysics and Meteorology, University of Cologne and Treve Nicol of the Department of Meteorology, University of Reading for useful discussions. The authors also acknowledge the support of Departmental technical staff.

¹J. G. C. Ball, B. E. Reed, R. G. Grainger, D. M. Peters, T. A. Mather, and D. M. Pyle, *J. Geophys. Res.: Atmos.* **120**, 7747, doi:10.1002/2015jd023521 (2015).

²P. N. Francis, M. C. Cooke, and R. W. Saunders, *J. Geophys. Res.: Atmos.* **117**, D00U09, doi:10.1029/2011JD016788 (2012).

³G. E. Hill and D. S. Woffinden, *J. Appl. Meteorol.* **19**, 1285 (1980).

⁴G. E. Hill, *J. Atmos. Oceanic Technol.* **6**, 961 (1989).

⁵R. G. Harrison, K. A. Nicoll, and A. G. Lomas, *Rev. Sci. Instrum.* **83**, 036106 (2012).

⁶R. V. L. Hartley, *Proc. IRE* **30**, 144 (1942).

⁷R. N. Bracewell, *The Hartley Transform* (Oxford University Press, New York, 1986).

⁸E. P. Lozowski, J. R. Stallabass, and P. F. Hearty, *J. Clim. Appl. Meteorol.* **22**, 2053 (1983).